

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX061220\
 Data File : VX016719.D
 Acq On : 12 Jun 2020 11:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 13 04:41:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	50.000	40.787	18.4	63	0.00
3 P	Chloromethane	50.000	36.843	26.3#	61	0.00
4 C	Vinyl Chloride	50.000	40.480	19.0#	66	0.00
5 T	Bromomethane	50.000	44.857	10.3	76	0.00
6 T	Chloroethane	50.000	45.527	8.9	73	0.00
7 T	Trichlorofluoromethane	50.000	51.395	-2.8	82	0.00
8 T	Diethyl Ether	50.000	46.419	7.2	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.699	-5.4	85	0.00
10 T	Methyl Iodide	50.000	44.973	10.1	75	0.00
11 T	Tert butyl alcohol	250.000	208.472	16.6	67	0.00
12 CM	1,1-Dichloroethene	50.000	47.780	4.4#	78	0.00
13 T	Acrolein	250.000	180.580	27.8#	62	0.00
14 T	Allyl chloride	50.000	40.933	18.1	65	0.00
15 T	Acrylonitrile	250.000	210.764	15.7	66	0.00
16 T	Acetone	250.000	228.986	8.4	72	0.00
17 T	Carbon Disulfide	50.000	39.995	20.0	65	0.00
18 T	Methyl Acetate	50.000	45.761	8.5	72	0.00
19 T	Methyl tert-butyl Ether	50.000	47.224	5.6	75	0.00
20 T	Methylene Chloride	50.000	45.085	9.8	74	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.118	9.8	73	0.00
22 T	Diisopropyl ether	50.000	42.077	15.8	66	0.00
23 T	Vinyl Acetate	250.000	210.314	15.9	64	0.00
24 P	1,1-Dichloroethane	50.000	44.874	10.3	71	0.00
25 T	2-Butanone	250.000	199.220	20.3	62	0.00
26 T	2,2-Dichloropropane	50.000	53.984	-8.0	89	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.787	8.4	74	0.00
28 T	Bromochloromethane	50.000	39.737	20.5	68	0.00
29 T	Tetrahydrofuran	250.000	196.234	21.5	60	0.00
30 C	Chloroform	50.000	48.304	3.4#	77	0.00
31 T	Cyclohexane	50.000	39.509	21.0	64	0.00
32 T	1,1,1-Trichloroethane	50.000	49.167	1.7	78	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.025	8.0	76	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	75	0.00
35 S	Dibromofluoromethane	50.000	54.144	-8.3	81	0.00
36 T	1,1-Dichloropropene	50.000	49.938	0.1	73	0.00
37 T	Ethyl Acetate	50.000	43.335	13.3	63	0.00
38 T	Carbon Tetrachloride	50.000	55.466	-10.9	80	0.00
39 T	Methylcyclohexane	50.000	51.414	-2.8	74	0.00
40 TM	Benzene	50.000	49.636	0.7	72	0.00
41 T	Methacrylonitrile	50.000	43.300	13.4	64	0.00
42 TM	1,2-Dichloroethane	50.000	52.013	-4.0	75	0.00
43 T	Isopropyl Acetate	50.000	44.385	11.2	64	0.00
44 TM	Trichloroethene	50.000	47.780	4.4	71	0.00
45 C	1,2-Dichloropropane	50.000	49.196	1.6#	71	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
 Data File : VX016719.D
 Acq On : 12 Jun 2020 11:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 13 04:41:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Dibromomethane	50.000	51.901	-3.8	75	0.00
47 T	Bromodichloromethane	50.000	54.473	-8.9	78	0.00
48 T	Methyl methacrylate	50.000	45.242	9.5	64	0.00
49 T	1,4-Dioxane	1000.000	975.634	2.4	69	0.00
50 S	Toluene-d8	50.000	52.151	-4.3	77	0.00
51 T	4-Methyl-2-Pentanone	250.000	228.937	8.4	64	0.00
52 CM	Toluene	50.000	52.321	-4.6#	75	0.00
53 T	t-1,3-Dichloropropene	50.000	53.552	-7.1	76	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.591	-5.2	75	0.00
55 T	1,1,2-Trichloroethane	50.000	54.594	-9.2	79	0.00
56 T	Ethyl methacrylate	50.000	51.000	-2.0	71	0.00
57 T	1,3-Dichloropropane	50.000	51.920	-3.8	74	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.622	2.2	71	0.00
59 T	2-Hexanone	250.000	232.054	7.2	64	0.00
60 T	Dibromochloromethane	50.000	58.318	-16.6	83	0.00
61 T	1,2-Dibromoethane	50.000	54.312	-8.6	78	0.00
62 S	4-Bromofluorobenzene	50.000	52.747	-5.5	80	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	76	0.00
64 T	Tetrachloroethene	50.000	43.181	13.6	65	0.00
65 PM	Chlorobenzene	50.000	53.224	-6.4	79	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.112	-12.2	83	0.00
67 C	Ethyl Benzene	50.000	51.524	-3.0#	75	0.00
68 T	m/p-Xylenes	100.000	106.427	-6.4	76	0.00
69 T	o-Xylene	50.000	52.673	-5.3	77	0.00
70 T	Styrene	50.000	55.101	-10.2	78	0.00
71 P	Bromoform	50.000	58.164	-16.3	83	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	80	0.00
73 T	Isopropylbenzene	50.000	50.498	-1.0	78	0.00
74 T	N-amyl acetate	50.000	41.771	16.5	64	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	58.830	-17.7	89	0.00
76 T	1,2,3-Trichloropropane	50.000	48.548	2.9	76	0.00
77 T	Bromobenzene	50.000	51.657	-3.3	82	0.00
78 T	n-propylbenzene	50.000	50.915	-1.8	78	0.00
79 T	2-Chlorotoluene	50.000	50.407	-0.8	78	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.211	-4.4	80	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.704	4.6	75	0.00
82 T	4-Chlorotoluene	50.000	50.726	-1.5	79	0.00
83 T	tert-Butylbenzene	50.000	57.203	-14.4	87	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.017	-6.0	81	0.00
85 T	sec-Butylbenzene	50.000	53.387	-6.8	82	0.00
86 T	p-Isopropyltoluene	50.000	55.469	-10.9	85	0.00
87 T	1,3-Dichlorobenzene	50.000	53.130	-6.3	84	0.00
88 T	1,4-Dichlorobenzene	50.000	52.221	-4.4	83	0.00
89 T	n-Butylbenzene	50.000	56.128	-12.3	85	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
Data File : VX016719.D
Acq On : 12 Jun 2020 11:23
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 13 04:41:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	50.000	55.716	-11.4	86	0.00
91 T	1,2-Dichlorobenzene	50.000	53.147	-6.3	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.699	2.6	76	0.00
93 T	1,2,4-Trichlorobenzene	50.000	58.112	-16.2	90	0.00
94 T	Hexachlorobutadiene	50.000	67.781	-35.6#	110	0.00
95 T	Naphthalene	50.000	52.735	-5.5	81	0.00
96 T	1,2,3-Trichlorobenzene	50.000	57.627	-15.3	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6