

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
 Data File : VX036117.D
 Acq On : 12 Jun 2023 12:48
 Operator : JC/MD
 Sample : VX0612MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0612MBL01

Quant Time: Jun 13 04:46:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	227615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	363425	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150921	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166480	45.042	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.080%
35) Dibromofluoromethane	5.385	113	112673	46.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	8.647	98	430876	49.116	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.079	95	168483	48.651	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
Data File : VX036117.D
Acq On : 12 Jun 2023 12:48
Operator : JC/MD
Sample : VX0612MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0612MBL01

Quant Time: Jun 13 04:46:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

