

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
 Data File : VX036122.D
 Acq On : 12 Jun 2023 14:51
 Operator : JC/MD
 Sample : PB153331ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153331ZHE#02

Quant Time: Jun 13 04:49:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	210869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	346768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159776	46.661	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.320%
35) Dibromofluoromethane	5.385	113	106881	45.826	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.660%
50) Toluene-d8	8.647	98	417822	49.915	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.840%
62) 4-Bromofluorobenzene	11.079	95	161010	48.726	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.460%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	30939	21.901	ug/l	95
20) Methylene Chloride	2.788	84	5852	2.151	ug/l	95
37) Ethyl Acetate	4.721	43	53267	15.380	ug/l	99
43) Isopropyl Acetate	6.336	43	209500	36.789	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
Data File : VX036122.D
Acq On : 12 Jun 2023 14:51
Operator : JC/MD
Sample : PB153331ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153331ZHE#02

Quant Time: Jun 13 04:49:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

