

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
 Data File : VX036126.D
 Acq On : 12 Jun 2023 16:25
 Operator : JC/MD
 Sample : PB153331ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153331ZHE#06

Quant Time: Jun 13 04:50:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	208112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	340762	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	312740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158033	46.763	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.520%
35) Dibromofluoromethane	5.385	113	106409	46.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	413343	50.251	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.079	95	158848	48.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.840%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	25250	18.111	ug/l	95
20) Methylene Chloride	2.788	84	3495	1.301	ug/l #	90
37) Ethyl Acetate	4.721	43	43022	12.641	ug/l	100
43) Isopropyl Acetate	6.342	43	151580	27.087	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
Data File : VX036126.D
Acq On : 12 Jun 2023 16:25
Operator : JC/MD
Sample : PB153331ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153331ZHE#06

Quant Time: Jun 13 04:50:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

