

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
 Data File : VX036127.D
 Acq On : 12 Jun 2023 16:49
 Operator : JC/MD
 Sample : PB153331ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153331ZHE#07

Quant Time: Jun 13 04:51:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	224248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	372004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	339040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	171119	46.992	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.980%
35) Dibromofluoromethane	5.385	113	115920	46.330	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	8.647	98	446432	49.715	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	172468	48.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.300%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	26517	17.651	ug/l	96
20) Methylene Chloride	2.788	84	5290	1.828	ug/l	96
37) Ethyl Acetate	4.721	43	52438	14.114	ug/l	99
43) Isopropyl Acetate	6.342	43	206217	33.756	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061223\
Data File : VX036127.D
Acq On : 12 Jun 2023 16:49
Operator : JC/MD
Sample : PB153331ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153331ZHE#07

Quant Time: Jun 13 04:51:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

