

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
 Data File : VX041831.D
 Acq On : 12 Jun 2024 10:53
 Operator : JC/MD
 Sample : VX0612MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0612MBL01

Quant Time: Jun 13 03:38:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	176063	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	275554	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137681	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92888	43.017	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.040%
35) Dibromofluoromethane	5.379	113	84260	44.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.460%
50) Toluene-d8	8.647	98	326303	49.017	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.040%
62) 4-Bromofluorobenzene	11.079	95	124449	50.652	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
Data File : VX041831.D
Acq On : 12 Jun 2024 10:53
Operator : JC/MD
Sample : VX0612MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0612MBL01

Quant Time: Jun 13 03:38:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

