

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
 Data File : VX041835.D
 Acq On : 12 Jun 2024 12:29
 Operator : JC/MD
 Sample : PB161443TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161443TB

Quant Time: Jun 13 03:41:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169229	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	267233	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	135517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89915	43.321	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.640%
35) Dibromofluoromethane	5.385	113	81264	43.983	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.960%
50) Toluene-d8	8.647	98	323032	50.037	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.080%
62) 4-Bromofluorobenzene	11.079	95	123370	51.777	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.560%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	89219	96.179	ug/l	91
18) Methyl Acetate	2.709	43	2228	1.124	ug/l	96
20) Methylene Chloride	2.788	84	6643	3.577	ug/l	91
25) 2-Butanone	4.574	43	22124	17.151	ug/l	93
43) Isopropyl Acetate	6.342	43	127760	29.971	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
Data File : VX041835.D
Acq On : 12 Jun 2024 12:29
Operator : JC/MD
Sample : PB161443TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161443TB

Quant Time: Jun 13 03:41:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

