

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
 Data File : VX041836.D
 Acq On : 12 Jun 2024 12:52
 Operator : JC/MD
 Sample : PB161443ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161443ZHE#01

Quant Time: Jun 13 03:42:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	177312	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	282042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	148480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93539	43.013	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.020%
35) Dibromofluoromethane	5.385	113	87331	44.785	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.560%
50) Toluene-d8	8.647	98	337195	49.488	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.079	95	131554	52.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.620%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	91867	94.454	ug/l	94
18) Methyl Acetate	2.709	43	3827	1.843	ug/l	95
20) Methylene Chloride	2.788	84	7219	3.709	ug/l	91
25) 2-Butanone	4.568	43	22727	16.815	ug/l	92
43) Isopropyl Acetate	6.342	43	130817	29.077	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
Data File : VX041836.D
Acq On : 12 Jun 2024 12:52
Operator : JC/MD
Sample : PB161443ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161443ZHE#01

Quant Time: Jun 13 03:42:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

