

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
 Data File : VX041837.D
 Acq On : 12 Jun 2024 13:15
 Operator : JC/MD
 Sample : P2784-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WCS-TPD

Quant Time: Jun 13 03:42:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169648	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	272609	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91179	43.822	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.640%
35) Dibromofluoromethane	5.385	113	83278	44.184	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.360%
50) Toluene-d8	8.647	98	325605	49.441	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	124269	51.125	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.260%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	91758	98.772	ug/l	90
18) Methyl Acetate	2.709	43	3979	2.003	ug/l	99
20) Methylene Chloride	2.788	84	7985	4.288	ug/l	# 92
25) 2-Butanone	4.574	43	22431	17.346	ug/l	99
37) Ethyl Acetate	4.739	43	3455	1.200	ug/l	# 92
43) Isopropyl Acetate	6.342	43	124236	28.570	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061224\
Data File : VX041837.D
Acq On : 12 Jun 2024 13:15
Operator : JC/MD
Sample : P2784-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WCS-TPD

Quant Time: Jun 13 03:42:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

