

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
 Data File : VX046658.D
 Acq On : 12 Jun 2025 12:48
 Operator : JC/MD
 Sample : VX0612MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0612MBL01

Quant Time: Jun 13 01:41:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	104695	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50000	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56116	52.192	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.380%
35) Dibromofluoromethane	5.397	113	45477	59.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.140%
50) Toluene-d8	8.653	98	149585	58.930	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	117.860%#
62) 4-Bromofluorobenzene	11.079	95	63161	60.140	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	120.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
Data File : VX046658.D
Acq On : 12 Jun 2025 12:48
Operator : JC/MD
Sample : VX0612MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0612MBL01

Quant Time: Jun 13 01:41:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

