

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061225\  
 Data File : VX046659.D  
 Acq On : 12 Jun 2025 13:10  
 Operator : JC/MD  
 Sample : VX0612WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0612WBL01

Quant Time: Jun 13 01:42:06 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060625W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 06 16:56:12 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.568  | 168  | 92224    | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.775  | 114  | 180952   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 188641   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 95358    | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 5.971  | 65    | 75758    | 46.173   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =    | 92.340%  |
| 35) Dibromofluoromethane    | 5.404  | 113   | 64522    | 48.487   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 96.980%  |
| 50) Toluene-d8              | 8.653  | 98    | 229993   | 52.424   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =    | 104.840% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 99118    | 54.605   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =    | 109.200% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061225\  
Data File : VX046659.D  
Acq On : 12 Jun 2025 13:10  
Operator : JC/MD  
Sample : VX0612WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0612WBL01

Quant Time: Jun 13 01:42:06 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X060625W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 06 16:56:12 2025  
Response via : Initial Calibration

