

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
 Data File : VX046661.D
 Acq On : 12 Jun 2025 14:04
 Operator : JC/MD
 Sample : Q2249-01RE 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-6.5-060525RE

Quant Time: Jun 12 14:23:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130843	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.769	114	256854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	135069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106796	45.878	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.760%
35) Dibromofluoromethane	5.391	113	91952	48.680	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	8.653	98	329160	52.856	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.720%
62) 4-Bromofluorobenzene	11.079	95	138615	53.798	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.600%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.343	101	19559	11.384	ug/l	94
16) Acetone	2.392	43	1345	6.039	ug/l #	83
27) cis-1,2-Dichloroethene	4.513	96	2540	1.255	ug/l	87
44) Trichloroethene	7.129	130	197730	102.453	ug/l	99
64) Tetrachloroethene	9.275	164	1427	0.810	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
Data File : VX046661.D
Acq On : 12 Jun 2025 14:04
Operator : JC/MD
Sample : Q2249-01RE 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-6.5-060525RE

Quant Time: Jun 12 14:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

