

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
 Data File : VX046666.D
 Acq On : 12 Jun 2025 16:10
 Operator : JC/MD
 Sample : Q2275-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-060925

Quant Time: Jun 13 01:45:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	56691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	96683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58286	57.790	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.580%	
35) Dibromofluoromethane	5.397	113	47119	66.271	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 132.540%#	
50) Toluene-d8	8.653	98	149280	63.684	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 127.360%#	
62) 4-Bromofluorobenzene	11.079	95	62307	64.244	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 128.480%#	
Target Compounds						
16) Acetone	2.398	43	1332	7.896	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061225\
Data File : VX046666.D
Acq On : 12 Jun 2025 16:10
Operator : JC/MD
Sample : Q2275-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-060925

Quant Time: Jun 13 01:45:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

