

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061319\
 Data File : VX010283.D
 Acq On : 13 Jun 2019 19:19
 Operator : JC/SP
 Sample : K3309-01DL 1000X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW009-190610DL

Quant Time: Jun 14 04:23:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	328148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	497981	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	450417	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212124	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	146881	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
35) Dibromofluoromethane	5.50	113	151177	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	8.71	98	582004	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	198938	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
Target Compounds						
4) Vinyl Chloride	1.41	62	80068	28.410	ug/l	96
21) trans-1,2-Dichloroethene	3.16	96	1380	0.423	ug/l #	58
27) cis-1,2-Dichloroethene	4.60	96	78365	21.596	ug/l	69
91) 1,2-Dichlorobenzene	12.39	146	14601	2.094	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061319\
Data File : VX010283.D
Acq On : 13 Jun 2019 19:19
Operator : JC/SP
Sample : K3309-01DL 1000X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW009-190610DL

Quant Time: Jun 14 04:23:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

