

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061319\
 Data File : VX010278.D
 Acq On : 13 Jun 2019 17:22
 Operator : JC/SP
 Sample : K3341-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-6-12-19

Quant Time: Jun 14 09:31:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	325960	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	490206	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	437070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208125	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	145767	45.81	ug/l	0.00
Spiked Amount			Recovery	=	91.62%	
35) Dibromofluoromethane	5.50	113	147580	48.81	ug/l	0.00
Spiked Amount			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	568672	50.18	ug/l	0.00
Spiked Amount			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	11.13	95	194575	45.98	ug/l	0.00
Spiked Amount			Recovery	=	91.96%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061319\
Data File : VX010278.D
Acq On : 13 Jun 2019 17:22
Operator : JC/SP
Sample : K3341-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-6-12-19

Quant Time: Jun 14 09:31:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22

Instrument :

MSVOA_X

ClientSampled :


FB-6-12-19

Tot Ion:168 Resp: 325960

Ion Ratio Lower Upper

168 100

99 46.2 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



5.66

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 45.814 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010278.D


Acq: 13 Jun 2019 17:22

Tgt Ion: 65 Resp: 145767

Ion Ratio Lower Upper

65 100

67 55.2 0.0 108.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

60000

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22

Instrument :

MSVOA_X

ClientSampled :

FB-6-12-19


Tot Ion:114 Resp: 490206

Ion Ratio Lower Upper

114 100

63 16.1 0.0 47.4




88 14.6 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.808 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22


Tgt Ion:113 Resp: 147580

Ion Ratio Lower Upper

113 100

111 102.5 82.2 123.2



192 23.2 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.181 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22

Instrument :

MSVOA_X

ClientSampleId :





FB-6-12-19

Tot Ion: 98 Resp: 568672

Ion Ratio Lower Upper

98 100

100 64.3 51.5 77.3



#62

4-Bromofluorobenzene

Concen: 45.975 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22




Tgt Ion: 95 Resp: 194575


Ion Ratio Lower Upper

95 100

174 94.4 0.0 160.4

176 92.4 0.0 154.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22

Instrument :

MSVOA_X

ClientSampled :

FB-6-12-19


Tot Ion:117 Resp: 437070

Ion Ratio Lower Upper

117 100

82 51.0 49.2 73.8




119 32.4 25.2 37.8



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX010278.D

Acq: 13 Jun 2019 17:22


Tgt Ion:152 Resp: 208125

Ion Ratio Lower Upper

152 100

115 53.6 41.4 124.2

150 155.6 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

