

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036132.D
 Acq On : 13 Jun 2023 09:38
 Operator : JC/MD
 Sample : VX0613MBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0613MBL01

Quant Time: Jun 14 00:50:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	226517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	372320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	333724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	156825	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	175288	47.655	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.300%
35) Dibromofluoromethane	5.379	113	118037	47.136	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	8.647	98	441902	49.169	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.340%
62) 4-Bromofluorobenzene	11.079	95	174403	49.157	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036132.D
Acq On : 13 Jun 2023 09:38
Operator : JC/MD
Sample : VX0613MBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0613MBL01

Quant Time: Jun 14 00:50:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

