

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036136.D
 Acq On : 13 Jun 2023 12:08
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jun 14 00:52:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	213918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	352423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	309797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147314	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	164423	47.333	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.660%
35) Dibromofluoromethane	5.379	113	110297	46.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.647	98	414342	48.705	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.420%
62) 4-Bromofluorobenzene	11.079	95	162149	48.284	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036136.D
Acq On : 13 Jun 2023 12:08
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jun 14 00:52:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

