

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036137.D
 Acq On : 13 Jun 2023 12:31
 Operator : JC/MD
 Sample : 03178-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jun 14 00:53:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	202924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	333723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	291277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	131929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	155132	47.078	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.160%
35) Dibromofluoromethane	5.385	113	105034	46.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.580%
50) Toluene-d8	8.647	98	395406	49.084	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	150284	47.258	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036137.D
Acq On : 13 Jun 2023 12:31
Operator : JC/MD
Sample : 03178-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jun 14 00:53:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

