

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036143.D
 Acq On : 13 Jun 2023 14:52
 Operator : JC/MD
 Sample : 03178-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jun 14 00:55:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	197299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	327806	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294152	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159382	49.747	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.500%		
35) Dibromofluoromethane	5.385	113	103135	46.777	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.560%		
50) Toluene-d8	8.647	98	386185	48.804	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.600%		
62) 4-Bromofluorobenzene	11.079	95	153797	49.236	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.480%		
Target Compounds						
16) Acetone	2.410	43	3705	2.803	ug/l	95
19) Methyl tert-butyl Ether	3.111	73	2330	0.301	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036143.D
Acq On : 13 Jun 2023 14:52
Operator : JC/MD
Sample : 03178-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jun 14 00:55:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

