

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036144.D
 Acq On : 13 Jun 2023 15:15
 Operator : JC/MD
 Sample : 03178-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2A

Quant Time: Jun 14 00:55:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/14/2023
 Supervised By : Mahesh Dadoda 06/14/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	201312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332732	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	298288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	158586	48.512	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.020%		
35) Dibromofluoromethane	5.385	113	105205	47.010	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.020%		
50) Toluene-d8	8.647	98	397826	49.531	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.060%		
62) 4-Bromofluorobenzene	11.079	95	155187	48.945	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.900%		
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	2201	1.632	ug/l	98
31) Cyclohexane	5.471	56	1103m	0.322	ug/l	
85) sec-Butylbenzene	11.890	105	2759	0.286	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	1555	0.342	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036144.D
Acq On : 13 Jun 2023 15:15
Operator : JC/MD
Sample : 03178-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2A

Quant Time: Jun 14 00:55:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/14/2023
Supervised By : Mahesh Dadoda 06/14/2023

