

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036145.D
 Acq On : 13 Jun 2023 15:39
 Operator : JC/MD
 Sample : 03178-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jun 14 00:55:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	211140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	347634	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	310109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160586	46.837	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.680%		
35) Dibromofluoromethane	5.385	113	106794	45.674	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.340%		
50) Toluene-d8	8.647	98	415331	49.494	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.980%		
62) 4-Bromofluorobenzene	11.079	95	158756	47.925	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.840%		
Target Compounds					Qvalue	
16) Acetone	2.398	43	1908	1.349	ug/l #	79
20) Methylene Chloride	2.782	84	909	0.334	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036145.D
Acq On : 13 Jun 2023 15:39
Operator : JC/MD
Sample : 03178-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jun 14 00:55:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

