

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
 Data File : VX036151.D
 Acq On : 13 Jun 2023 18:00
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 14 00:58:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	216127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	353182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	313637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	164373	46.835	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.680%
35) Dibromofluoromethane	5.385	113	109641	46.155	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	422641	49.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.140%
62) 4-Bromofluorobenzene	11.079	95	159081	47.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.540%
Target Compounds						
					Qvalue	
16) Acetone	2.410	43	1489	1.028	ug/l #	76
20) Methylene Chloride	2.794	84	1958	0.702	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061323\
Data File : VX036151.D
Acq On : 13 Jun 2023 18:00
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 14 00:58:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

