

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041844.D
 Acq On : 13 Jun 2024 11:25
 Operator : JC/MD
 Sample : PB161467TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467TB

Quant Time: Jun 14 00:54:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	166828	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	261909	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87361	42.697	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.400%
35) Dibromofluoromethane	5.385	113	79676	44.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.000%
50) Toluene-d8	8.647	98	310542	49.080	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	116499	49.887	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.780%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	86298	94.298	ug/l	90
18) Methyl Acetate	2.709	43	2603	1.332	ug/l	96
20) Methylene Chloride	2.788	84	9505	5.191	ug/l	88
25) 2-Butanone	4.574	43	21488	16.898	ug/l	98
43) Isopropyl Acetate	6.342	43	121596	29.105	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041844.D
Acq On : 13 Jun 2024 11:25
Operator : JC/MD
Sample : PB161467TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467TB

Quant Time: Jun 14 00:54:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

