

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041847.D
 Acq On : 13 Jun 2024 12:35
 Operator : JC/MD
 Sample : PB161467ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#02

Quant Time: Jun 14 00:56:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	162354	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	258784	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	266739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85893	43.136	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.280%
35) Dibromofluoromethane	5.385	113	79032	44.171	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.340%
50) Toluene-d8	8.647	98	311081	49.759	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	120430	52.193	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.380%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	90108	101.456	ug/l	# 89
18) Methyl Acetate	2.709	43	4212	2.215	ug/l	95
20) Methylene Chloride	2.788	84	9023	5.064	ug/l	89
25) 2-Butanone	4.574	43	23141	18.699	ug/l	98
43) Isopropyl Acetate	6.342	43	126246	30.583	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041847.D
Acq On : 13 Jun 2024 12:35
Operator : JC/MD
Sample : PB161467ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#02

Quant Time: Jun 14 00:56:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

