

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041848.D
 Acq On : 13 Jun 2024 12:58
 Operator : JC/MD
 Sample : PB161467ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#03

Quant Time: Jun 14 00:56:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	177451	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	271642	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89674	41.203	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.400%
35) Dibromofluoromethane	5.385	113	83828	44.634	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.260%
50) Toluene-d8	8.647	98	325715	49.634	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.260%
62) 4-Bromofluorobenzene	11.079	95	121923	50.339	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.680%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	91441	93.922	ug/l	90
18) Methyl Acetate	2.709	43	5218	2.511	ug/l	99
20) Methylene Chloride	2.788	84	9105	4.675	ug/l	92
25) 2-Butanone	4.574	43	22178	16.396	ug/l	100
37) Ethyl Acetate	4.745	43	3413	1.189	ug/l	# 86
43) Isopropyl Acetate	6.348	43	124148	28.651	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041848.D
Acq On : 13 Jun 2024 12:58
Operator : JC/MD
Sample : PB161467ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#03

Quant Time: Jun 14 00:56:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

