

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041849.D
 Acq On : 13 Jun 2024 13:21
 Operator : JC/MD
 Sample : PB161467ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#04

Quant Time: Jun 14 00:57:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	156472	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	241705	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80056	41.716	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.440%
35) Dibromofluoromethane	5.385	113	73425	43.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.880%
50) Toluene-d8	8.647	98	290488	49.748	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.500%
62) 4-Bromofluorobenzene	11.079	95	105810	49.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.200%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	88939	104.000	ug/l	91
18) Methyl Acetate	2.715	43	6570	3.585	ug/l	100
20) Methylene Chloride	2.788	84	9431	5.491	ug/l	86
25) 2-Butanone	4.580	43	20917	17.537	ug/l	93
37) Ethyl Acetate	4.745	43	3008	1.178	ug/l	# 92
43) Isopropyl Acetate	6.342	43	122179	31.689	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041849.D
Acq On : 13 Jun 2024 13:21
Operator : JC/MD
Sample : PB161467ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#04

Quant Time: Jun 14 00:57:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

