

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041851.D
 Acq On : 13 Jun 2024 14:08
 Operator : JC/MD
 Sample : PB161467ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#06

Quant Time: Jun 14 00:58:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	168297	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	268019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91022	44.098	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.200%
35) Dibromofluoromethane	5.385	113	82164	44.339	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.680%
50) Toluene-d8	8.647	98	320104	49.438	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	122725	51.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.700%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	88854	96.322	ug/l	# 88
18) Methyl Acetate	2.709	43	6970	3.536	ug/l	96
20) Methylene Chloride	2.788	84	9767	5.288	ug/l	95
25) 2-Butanone	4.574	43	21504	16.763	ug/l	93
43) Isopropyl Acetate	6.348	43	119888	28.042	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041851.D
Acq On : 13 Jun 2024 14:08
Operator : JC/MD
Sample : PB161467ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#06

Quant Time: Jun 14 00:58:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

