

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041857.D
 Acq On : 13 Jun 2024 16:28
 Operator : JC/MD
 Sample : PB161467ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#12

Quant Time: Jun 14 01:00:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158538	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	251027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82049	42.197	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	84.400%	
35) Dibromofluoromethane	5.385	113	76659	44.169	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	88.340%	
50) Toluene-d8	8.647	98	300306	49.520	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.040%	
62) 4-Bromofluorobenzene	11.079	95	111187	49.676	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.360%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	90983	105.042	ug/l	# 89
18) Methyl Acetate	2.709	43	6924	3.729	ug/l	97
20) Methylene Chloride	2.782	84	10703	6.151	ug/l	88
25) 2-Butanone	4.574	43	22046	18.243	ug/l	100
37) Ethyl Acetate	4.757	43	3614	1.363	ug/l	# 84
43) Isopropyl Acetate	6.348	43	121424	30.324	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041857.D
Acq On : 13 Jun 2024 16:28
Operator : JC/MD
Sample : PB161467ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#12

Quant Time: Jun 14 01:00:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

