

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041859.D
 Acq On : 13 Jun 2024 17:15
 Operator : JC/MD
 Sample : PB161467ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161467ZHE#14

Quant Time: Jun 14 01:50:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159029	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	247002	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81259	41.662	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.320%
35) Dibromofluoromethane	5.385	113	75832	44.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.800%
50) Toluene-d8	8.646	98	292474	49.014	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.020%
62) 4-Bromofluorobenzene	11.079	95	105983	48.123	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.240%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	87383	100.405	ug/l	90
18) Methyl Acetate	2.715	43	7163	3.846	ug/l	97
20) Methylene Chloride	2.782	84	9907	5.676	ug/l	90
25) 2-Butanone	4.580	43	22190	18.305	ug/l	93
43) Isopropyl Acetate	6.348	43	122947	31.205	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041859.D
Acq On : 13 Jun 2024 17:15
Operator : JC/MD
Sample : PB161467ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161467ZHE#14

Quant Time: Jun 14 01:50:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

