

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041862.D
 Acq On : 13 Jun 2024 18:25
 Operator : JC/MD
 Sample : P2853-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2

Quant Time: Jun 14 01:51:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	171965	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	265976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125453	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89161	42.275	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.540%
35) Dibromofluoromethane	5.391	113	83848	45.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.200%
50) Toluene-d8	8.647	98	317780	49.456	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	115539	48.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.440%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	18558	17.371	ug/l	# 84
18) Methyl Acetate	2.709	43	7849	3.898	ug/l	98
20) Methylene Chloride	2.788	84	10707	5.673	ug/l	# 89
25) 2-Butanone	4.580	43	10335	7.884	ug/l	94
37) Ethyl Acetate	4.751	43	3487	1.241	ug/l	# 90
43) Isopropyl Acetate	6.342	43	123170	29.031	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041862.D
Acq On : 13 Jun 2024 18:25
Operator : JC/MD
Sample : P2853-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2

Quant Time: Jun 14 01:51:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

