

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
 Data File : VX041864.D
 Acq On : 13 Jun 2024 19:12
 Operator : JC/MD
 Sample : P2853-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 4

Quant Time: Jun 14 01:52:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154279	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	236657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241126	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76638	40.502	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.000%
35) Dibromofluoromethane	5.391	113	74203	45.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.700%
50) Toluene-d8	8.647	98	285759	49.982	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.960%
62) 4-Bromofluorobenzene	11.079	95	105157	49.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	19330	20.609	ug/l	# 89
18) Methyl Acetate	2.715	43	7177	3.972	ug/l	99
20) Methylene Chloride	2.788	84	11375	6.718	ug/l	89
25) 2-Butanone	4.581	43	10734	9.127	ug/l	100
37) Ethyl Acetate	4.745	43	3339	1.335	ug/l	# 74
43) Isopropyl Acetate	6.348	43	123660	32.758	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061324\
Data File : VX041864.D
Acq On : 13 Jun 2024 19:12
Operator : JC/MD
Sample : P2853-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
4

Quant Time: Jun 14 01:52:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

