



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX061324\  
Data File : VX041839.D  
Acq On : 13 Jun 2024 08:21  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X052924W.M  
Title : SW846 8260  
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 15.9         | 13294      | PASS                |
| 75             | 95              | 30              | 60              | 45.3         | 37901      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 83592      | PASS                |
| 96             | 95              | 5               | 9               | 6.6          | 5534       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 949        | PASS                |
| 174            | 95              | 50              | 100             | 93.8         | 78427      | PASS                |
| 175            | 174             | 5               | 9               | 7.4          | 5799       | PASS                |
| 176            | 174             | 95              | 101             | 97.6         | 76509      | PASS                |
| 177            | 176             | 5               | 9               | 6.7          | 5150       | PASS                |