

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046678.D
 Acq On : 13 Jun 2025 10:47
 Operator : JC/MD
 Sample : VX0613WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0613WBL01

Quant Time: Jun 13 23:02:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061025W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jun 11 06:20:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.916	128	18213	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	98625	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	90050	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.970	65	42782	29.019	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.733%
60) 4-Bromofluorobenzene	11.079	95	43724	28.807	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.033%
63) Toluene-d8	8.653	98	117534	29.221	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046678.D
Acq On : 13 Jun 2025 10:47
Operator : JC/MD
Sample : VX0613WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0613WBL01

Quant Time: Jun 13 23:02:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061025W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jun 11 06:20:28 2025
Response via : Initial Calibration

