

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046683.D
 Acq On : 13 Jun 2025 12:58
 Operator : JC/MD
 Sample : VX0613MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0613MBL01

Quant Time: Jun 13 23:26:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	96124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189261	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.055	117	198520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	97830	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	81082	47.412	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.820%
35) Dibromofluoromethane	5.397	113	70322	50.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	8.647	98	245658	53.536	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.080%
62) 4-Bromofluorobenzene	11.079	95	103145	54.329	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046683.D
Acq On : 13 Jun 2025 12:58
Operator : JC/MD
Sample : VX0613MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0613MBL01

Quant Time: Jun 13 23:26:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

