

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046688.D
 Acq On : 13 Jun 2025 14:53
 Operator : JC/MD
 Sample : Q2287-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-SLAB

Quant Time: Jun 13 23:30:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	93474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	167468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76233	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	68212	41.017	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.040%
35) Dibromofluoromethane	5.416	113	680	0.552	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	1.100%#
50) Toluene-d8	8.653	98	185606	45.713	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.420%
62) 4-Bromofluorobenzene	11.079	95	72729	43.293	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.580%
Target Compounds						
16) Acetone	2.392	43	37346	60.731	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046688.D
Acq On : 13 Jun 2025 14:53
Operator : JC/MD
Sample : Q2287-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-SLAB

Quant Time: Jun 13 23:30:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

