

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046692.D
 Acq On : 13 Jun 2025 16:19
 Operator : JC/MD
 Sample : Q2296-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2

Quant Time: Jun 13 23:31:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	92827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	170984	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	81838	49.554	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.100%
35) Dibromofluoromethane	5.403	113	60599	48.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.653	98	215508	51.986	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.980%
62) 4-Bromofluorobenzene	11.079	95	83089	48.443	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.880%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	8230	17.052	ug/l	93
18) Methyl Acetate	2.721	43	7710	3.919	ug/l #	90
20) Methylene Chloride	2.806	84	3570	2.694	ug/l #	89
37) Ethyl Acetate	4.751	43	5452	3.110	ug/l	98
43) Isopropyl Acetate	6.367	43	5025	1.644	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046692.D
Acq On : 13 Jun 2025 16:19
Operator : JC/MD
Sample : Q2296-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2

Quant Time: Jun 13 23:31:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

