

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046693.D
 Acq On : 13 Jun 2025 16:41
 Operator : JC/MD
 Sample : Q2296-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-3

Quant Time: Jun 13 23:32:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	92370	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	169535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153393	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	74297	45.211	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.420%
35) Dibromofluoromethane	5.397	113	55243	44.309	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.620%
50) Toluene-d8	8.653	98	197948	48.158	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.079	95	76084	44.738	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.480%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	11389	21.918	ug/l	91
18) Methyl Acetate	2.721	43	10762	5.498	ug/l	94
20) Methylene Chloride	2.806	84	4101	3.111	ug/l	92
37) Ethyl Acetate	4.745	43	5802	3.338	ug/l	99
43) Isopropyl Acetate	6.361	43	4920	1.624	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046693.D
Acq On : 13 Jun 2025 16:41
Operator : JC/MD
Sample : Q2296-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-3

Quant Time: Jun 13 23:32:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

