

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046694.D
 Acq On : 13 Jun 2025 17:02
 Operator : JC/MD
 Sample : Q2296-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4

Quant Time: Jun 13 23:32:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	94731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	174083	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	85407	50.676	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.360%
35) Dibromofluoromethane	5.403	113	63799	49.835	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.680%
50) Toluene-d8	8.653	98	227177	53.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.640%
62) 4-Bromofluorobenzene	11.079	95	85948	49.218	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.440%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9005	17.951	ug/l	98
18) Methyl Acetate	2.721	43	10311	5.136	ug/l	94
20) Methylene Chloride	2.806	84	4461	3.299	ug/l	86
37) Ethyl Acetate	4.745	43	5563	3.117	ug/l #	73
43) Isopropyl Acetate	6.354	43	4956	1.593	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046694.D
Acq On : 13 Jun 2025 17:02
Operator : JC/MD
Sample : Q2296-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4

Quant Time: Jun 13 23:32:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

