

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046695.D
 Acq On : 13 Jun 2025 17:24
 Operator : JC/MD
 Sample : Q2296-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-5

Quant Time: Jun 13 23:33:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	92543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	173187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	151643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	76977	46.754	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.500%
35) Dibromofluoromethane	5.404	113	56790	44.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.180%
50) Toluene-d8	8.653	98	201537	47.997	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.079	95	77111	44.386	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.780%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	11094	21.438	ug/l	96
18) Methyl Acetate	2.721	43	11014	5.616	ug/l	97
20) Methylene Chloride	2.813	84	4359	3.300	ug/l #	83
37) Ethyl Acetate	4.745	43	5756	3.242	ug/l #	90
43) Isopropyl Acetate	6.373	43	4779	1.544	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046695.D
Acq On : 13 Jun 2025 17:24
Operator : JC/MD
Sample : Q2296-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-5

Quant Time: Jun 13 23:33:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

