

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046697.D
 Acq On : 13 Jun 2025 18:07
 Operator : JC/MD
 Sample : Q2297-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3

Quant Time: Jun 13 23:34:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	89762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	166230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	151499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	78825	49.359	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.720%
35) Dibromofluoromethane	5.403	113	58710	48.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.653	98	208464	51.725	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.440%
62) 4-Bromofluorobenzene	11.079	95	78986	47.368	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.740%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	18423	33.432	ug/l	96
18) Methyl Acetate	2.721	43	11483	6.037	ug/l	98
20) Methylene Chloride	2.806	84	4132	3.225	ug/l	82
37) Ethyl Acetate	4.751	43	5943	3.487	ug/l #	91
43) Isopropyl Acetate	6.361	43	4848	1.632	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046697.D
Acq On : 13 Jun 2025 18:07
Operator : JC/MD
Sample : Q2297-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3

Quant Time: Jun 13 23:34:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

