

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046700.D
 Acq On : 13 Jun 2025 19:12
 Operator : JC/MD
 Sample : Q2288-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-329

Quant Time: Jun 13 23:35:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	93014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	166226	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	70322	42.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.000%
35) Dibromofluoromethane	5.403	113	52792	43.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.380%
50) Toluene-d8	8.653	98	189175	46.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.880%
62) 4-Bromofluorobenzene	11.079	95	71518	42.890	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	85.780%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	22974	39.299	ug/l	99
18) Methyl Acetate	2.721	43	6824	3.462	ug/l	97
20) Methylene Chloride	2.806	84	21335	16.070	ug/l	91
37) Ethyl Acetate	4.751	43	2750	1.614	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046700.D
Acq On : 13 Jun 2025 19:12
Operator : JC/MD
Sample : Q2288-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-329

Quant Time: Jun 13 23:35:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

