

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
 Data File : VX046701.D
 Acq On : 13 Jun 2025 19:34
 Operator : JC/MD
 Sample : Q2287-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-SLABRE

Quant Time: Jun 13 23:36:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	95835	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	168297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	82976	48.666	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.340%
35) Dibromofluoromethane	5.556	113	43	0.035	ug/l	0.15
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.060%#
50) Toluene-d8	8.653	98	217204	53.231	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.460%
62) 4-Bromofluorobenzene	11.079	95	85311	50.533	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.060%
Target Compounds						
16) Acetone	2.392	43	52950	82.226	ug/l	94
20) Methylene Chloride	2.806	84	3083	2.254	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061325\
Data File : VX046701.D
Acq On : 13 Jun 2025 19:34
Operator : JC/MD
Sample : Q2287-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-SLABRE

Quant Time: Jun 13 23:36:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

