

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010293.D
 Acq On : 14 Jun 2019 12:22
 Operator : JC/SP
 Sample : K3309-02DL2 2000X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF001MW001-190610DL2

Quant Time: Jun 15 01:37:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	309415	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	466376	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	417323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140014	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
35) Dibromofluoromethane	5.50	113	141586	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	8.71	98	543882	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	184072	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
Target Compounds						
4) Vinyl Chloride	1.41	62	135854	51.123	ug/l	100
51) 4-Methyl-2-Pentanone	8.64	43	8812	2.228	ug/l	88
52) Toluene	8.79	92	9584	1.239	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	2863	0.451	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010293.D
Acq On : 14 Jun 2019 12:22
Operator : JC/SP
Sample : K3309-02DL2 2000X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF001MW001-190610DL2

Quant Time: Jun 15 01:37:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

