

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010298.D
 Acq On : 14 Jun 2019 14:19
 Operator : JC/SP
 Sample : K3315-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20190611

Quant Time: Jun 15 01:44:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	330858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	501760	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	447964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213156	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151164	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
35) Dibromofluoromethane	5.50	113	151387	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	589305	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.13	95	199365	46.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.04%	
Target Compounds						
16) Acetone	2.45	43	10886	7.167	ug/l #	82
52) Toluene	8.79	92	12014	1.444	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010298.D
Acq On : 14 Jun 2019 14:19
Operator : JC/SP
Sample : K3315-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20190611

Quant Time: Jun 15 01:44:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

