

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010303.D
 Acq On : 14 Jun 2019 16:16
 Operator : JC/SP
 Sample : K3315-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20190607

Quant Time: Jun 15 02:04:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	313507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	480981	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	428950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	202949	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	142015	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
35) Dibromofluoromethane	5.50	113	145182	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	8.71	98	558409	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.13	95	193420	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010303.D
Acq On : 14 Jun 2019 16:16
Operator : JC/SP
Sample : K3315-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20190607

Quant Time: Jun 15 02:04:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

