

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010304.D
 Acq On : 14 Jun 2019 16:39
 Operator : JC/SP
 Sample : K3315-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-20190607

Quant Time: Jun 15 02:08:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	314312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	479843	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	431511	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	141703	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
35) Dibromofluoromethane	5.50	113	144763	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	562992	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
62) 4-Bromofluorobenzene	11.13	95	190282	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	3710	1.342	ug/l	89
16) Acetone	2.45	43	3721	2.579	ug/l #	81
27) cis-1,2-Dichloroethene	4.60	96	2060	0.593	ug/l	70
44) Trichloroethene	7.21	130	182309	49.963	ug/l	88
49) 1,4-Dioxane	7.75	88	1081	11.904	ug/l #	79
64) Tetrachloroethene	9.34	164	7931	2.475	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010304.D
Acq On : 14 Jun 2019 16:39
Operator : JC/SP
Sample : K3315-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D1-20190607

Quant Time: Jun 15 02:08:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

