

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010308.D
 Acq On : 14 Jun 2019 18:12
 Operator : JC/SP
 Sample : K3384-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jun 15 02:55:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	294712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	449106	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	399936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	182047	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	135715	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
35) Dibromofluoromethane	5.50	113	136229	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
50) Toluene-d8	8.71	98	522623	50.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.13	95	176057	45.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010308.D
Acq On : 14 Jun 2019 18:12
Operator : JC/SP
Sample : K3384-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jun 15 02:55:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

