

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061419\
 Data File : VX010307.D
 Acq On : 14 Jun 2019 17:49
 Operator : JC/SP
 Sample : K3384-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 15 02:52:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	316603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	479709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	431398	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201137	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	145094	46.95	ug/l	0.00
Spiked Amount			Recovery	=	93.90%	
35) Dibromofluoromethane	5.50	113	146202	49.41	ug/l	0.00
Spiked Amount			Recovery	=	98.82%	
50) Toluene-d8	8.71	98	561744	50.65	ug/l	0.00
Spiked Amount			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	11.13	95	192247	46.42	ug/l	0.00
Spiked Amount			Recovery	=	92.84%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061419\
Data File : VX010307.D
Acq On : 14 Jun 2019 17:49
Operator : JC/SP
Sample : K3384-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 15 02:52:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX010307.D

Acq: 14 Jun 2019 17:49

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tot Ion:168 Resp: 316603

Ion Ratio Lower Upper

168 100

99 47.2 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

120000

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80 5.90



#33

1,2-Dichloroethane-d4

Concen: 46.950 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.01 min

Lab File: VX010307.D


Acq: 14 Jun 2019 17:49

Tgt Ion: 65 Resp: 145094

Ion Ratio Lower Upper

65 100

67 54.5 0.0 108.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000

50000

40000

30000


20000


10000

0

Time-->

5.90 6.00 6.10 6.20 6.30





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX010307.D

Acq: 14 Jun 2019 17:49

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4


Tot Ion:114 Resp: 479709

Ion Ratio Lower Upper

114 100

63 17.3 0.0 47.4




88 14.6 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.410 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX010307.D

Acq: 14 Jun 2019 17:49


Tgt Ion:113 Resp: 146202

Ion Ratio Lower Upper

113 100

111 100.9 82.2 123.2



192 22.6 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.655 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX010307.D
Acq: 14 Jun 2019 17:49




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tot Ion: 98 Resp: 561744
Ion Ratio Lower Upper
98 100
100 64.0 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 46.419 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX010307.D
Acq: 14 Jun 2019 17:49

Tgt Ion: 95 Resp: 192247
Ion Ratio Lower Upper
95 100
174 93.2 0.0 160.4
176 89.5 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010307.D
Acq: 14 Jun 2019 17:49

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tot Ion:117 Resp: 431398
Ion Ratio Lower Upper
117 100
82 50.6 49.2 73.8
119 31.3 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX010307.D
Acq: 14 Jun 2019 17:49

Tgt Ion:152 Resp: 201137
Ion Ratio Lower Upper
152 100
115 54.0 41.4 124.2
150 157.1 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

