

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061421\
 Data File : VX022691.D
 Acq On : 14 Jun 2021 13:53
 Operator : JC/MD
 Sample : M2688-03DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2DL

Quant Time: Jun 15 06:11:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67985	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	133337	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56299	46.045	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.100%		
35) Dibromofluoromethane	5.403	113	40638	47.769	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.540%		
50) Toluene-d8	8.653	98	164988	49.197	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.400%		
62) 4-Bromofluorobenzene	11.085	95	59077	46.578	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.160%		
Target Compounds						
					Qvalue	
31) Cyclohexane	5.476	56	10862	7.364	ug/l	# 81
39) Methylcyclohexane	7.385	83	3295	2.377	ug/l	86
40) Benzene	6.049	78	68903	17.433	ug/l	98
52) Toluene	8.726	92	171568	69.743	ug/l	99
67) Ethyl Benzene	10.195	91	51336	11.161	ug/l	99
68) m/p-Xylenes	10.305	106	66931	40.173	ug/l	93
69) o-Xylene	10.646	106	26545	16.444	ug/l	99
73) Isopropylbenzene	10.963	105	7289	1.826	ug/l	97
78) n-propylbenzene	11.305	91	17438	3.617	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	20237	6.086	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	100597	29.827	ug/l	97
95) Naphthalene	13.780	128	13003	4.529	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061421\
Data File : VX022691.D
Acq On : 14 Jun 2021 13:53
Operator : JC/MD
Sample : M2688-03DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2DL

Quant Time: Jun 15 06:11:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

