

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061421\
 Data File : VX022693.D
 Acq On : 14 Jun 2021 15:05
 Operator : JC/MD
 Sample : M2705-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRODUCT02-NMW-PS-D3

Quant Time: Jun 15 06:12:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52278	45.744	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.480%
35) Dibromofluoromethane	5.391	113	38462	48.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.653	98	154570	49.677	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.085	95	52494	44.609	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	9104	10.283	ug/l	96
6) Chloroethane	1.685	64	4640	8.655	ug/l	98
12) 1,1-Dichloroethene	2.319	96	1081	1.401	ug/l	92
16) Acetone	2.386	43	886	1.731	ug/l	91
24) 1,1-Dichloroethane	3.611	63	47482	27.721	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	120470	115.662	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	125726	86.084	ug/l	97
44) Trichloroethene	7.129	130	174152	202.297	ug/l	90
64) Tetrachloroethene	9.275	164	724	1.029	ug/l #	69
69) o-Xylene	10.646	106	875	0.584	ug/l	81
80) 1,3,5-Trimethylbenzene	11.451	105	2761	0.962	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061421\
Data File : VX022693.D
Acq On : 14 Jun 2021 15:05
Operator : JC/MD
Sample : M2705-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRODUCT02-NMW-PS-D3

Quant Time: Jun 15 06:12:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

